

SENSITECH INC.  
**CERTIFICATE OF CALIBRATION**



Sensitech Inc. • 800 Cummings Center • Suite 258X • Beverly, MA 01915-6197 • USA: 1-800-843-8367 • Int'l: +1-978-927-7033

**PRODUCT** TempTale®  
**STATEMENT OF CALIBRATION** Sensitech Inc. certifies that the device(s) on the Device List have been thoroughly tested and calibrated in accordance with Sensitech procedures and meet or exceed the accuracy specifications\*.  
 Calibration expires one year from the date of sale.  
 \*Specification details for Sensitech products are provided in the published datasheets.  
**REFERENCE EQUIPMENT\*\*** Fluke Hart Scientific thermometer model 1502A /Burns Engineering PRT model 5615 or equivalent  
 Temperature Accuracy: ±0.05°C; Calibration Range: -195°C to +232°C  
 \*\*Reference equipment calibration is accredited to ISO/IEC 17025 and is traceable through the National Institute of Standards and Technology (NIST), or other National Standards Institutes, to the International System of Units (SI). Calibration certificates for all reference equipment are retained by Sensitech Quality Assurance at our headquarters in Beverly, Massachusetts, USA.

**ATTESTED BY** Head of Quality

Daniel Steiner		02-Apr-2025
Name	Signature	Date

**SHIP TO** Carrier Transcold Hong Kong Ltd  
 7/F Chung Shun Knitting Centre  
 1-3 Wing Yip Street  
 Kwai Chung New Territories Hong Kong

**SOLD TO** Carrier Transcold Hong Kong Ltd  
 7/F Chung Shun Knitting Centre  
 1-3 Wing Yip Street  
 Kwai Chung New Territories Hong Kong

**ORDER NUMBER** HN3610171

**ORDER LINE NUMBER** 1

**ORDER RELEASE NUMBER** 1

**PACK ID** 850471

**RELEASED FROM** Hong Kong Facility

**RELEASED BY**

John Wong		4-Mar-2026
Name	Signature	Date

**DATE OF SALE** 2026-04-06




SENSITECH INC.  
**CERTIFICATE OF CALIBRATION**



Sensitech Inc. • 800 Cummings Center • Suite 258X • Beverly, MA 01915-6197 • USA: 1-800-843-8367 • Int'l: +1-978-927-7033

**PRODUCT** TempTale®  
**STATEMENT OF CALIBRATION** Sensitech Inc. certifies that the device(s) on the Device List have been thoroughly tested and calibrated in accordance with Sensitech procedures and meet or exceed the accuracy specifications\*.  
 Calibration expires one year from the date of sale.  
 \*Specification details for Sensitech products are provided in the published datasheets.  
**REFERENCE EQUIPMENT\*\*** Fluke Hart Scientific thermometer model 1502A /Burns Engineering PRT model 5615 or equivalent  
 Temperature Accuracy: ±0.05°C; Calibration Range: -195°C to +232°C  
 \*\*Reference equipment calibration is accredited to ISO/IEC 17025 and is traceable through the National Institute of Standards and Technology (NIST), or other National Standards Institutes, to the International System of Units (SI). Calibration certificates for all reference equipment are retained by Sensitech Quality Assurance at our headquarters in Beverly, Massachusetts, USA.

**ATTESTED BY** Head of Quality

	Daniel Steiner		02-Apr-2025
	Name	Signature	Date

**SHIP TO** Carrier Transcold Hong Kong Ltd  
 7/F Chung Shun Knitting Centre  
 1-3 Wing Yip Street  
 Kwai Chung New Territories Hong Kong

**SOLD TO** Carrier Transcold Hong Kong Ltd  
 7/F Chung Shun Knitting Centre  
 1-3 Wing Yip Street  
 Kwai Chung New Territories Hong Kong

**ORDER NUMBER** HN3610164


**ORDER LINE NUMBER** 9

**ORDER RELEASE NUMBER** 1

**PACK ID** 850254

**RELEASED FROM** Hong Kong Facility

**RELEASED BY**

	John Wong		2-Mar-2026
	Name	Signature	Date

**DATE OF SALE** 2026-04-06

# DEVICE LIST



<b>SHIP TO</b> Carrier Transicold Hong Kong Ltd 7/F Chung Shun Knitting Centre 1-3 Wing Yip Street Kwai Chung New Territories Hong Kong		<b>SOLD TO</b> Carrier Transicold Hong Kong Ltd 7/F Chung Shun Knitting Centre 1-3 Wing Yip Street Kwai Chung New Territories Hong Kong		<b>ORDER NUMBER</b> HN3610171	
				<b>ORDER LINE NUMBER</b> 1	
				<b>ORDER RELEASE NUMBER</b> 1	
				<b>PACK ID</b> 850471	
<b>Pack#:</b> 850471				<b>ORDER</b> HN3610171	
<b>BOX ID:</b> 26060B0661				<b>LINE: 1</b> <b>RELEASE 1</b>	
<b>SERIAL NUMBER</b>	<b>MODEL NUMBER</b>	<b>DESCRIPTION</b>		<b>CALIBRATION DATE</b>	
PMA2S51A30 <small>検体保管用冷凍庫 (-90 --70 設定)</small>	TUD32-02-094	TT ULTRA DI 3M PRB MU-NIHON A		4-Mar-2026	