

日常点検・警告アラームチェック表

20 26 年 6 月分

日常点検の結果、異常がなければ✓、異常ありの場合は「異常」と記載し、欄外に詳細を記載すること
 (日常点検項目: 冷蔵庫の異常音、異臭、その他外観上の異常、室内空調操作パネル異常の確認)

警告表示がなければ✓、警告表示ありの場合は「ALARM」と表記すること

日付 DD/MMM/YYYY	時間 hh:mm	日常 点検	室温 ROOM	恒温庫 INC-1	冷蔵庫 (左) REF-1 (LEFT)	冷蔵庫 (中) REF-2 (CENTER)	冷蔵庫(右) REF-3 (RIGHT)	REC 表示	署名
01/JUN/2026	08:27	✓	✓	✓	ALARM	✓	✓	✓	N.K
02/JUN/2026	09:55	✓	✓	✓	✓	✓	✓	✓	M.K
03/JUN/2026	09:07	✓	✓	✓	✓	✓	✓	✓	N.K
04/JUN/2026	09:56	✓	✓	✓	✓	✓	✓	✓	N.K
05/JUN/2026	11:00	✓	✓	✓	✓	✓	✓	✓	M.K
08/JUN/2026	10:44	✓	✓	✓	✓	✓	✓	✓	N.K
09/JUN/2026	11:33	✓	✓	✓	✓	✓	✓	✓	M.K
10/JUN/2026	11:39	✓	✓	✓	✓	✓	✓	✓	N.K
11/JUN/2026	10:06	✓	✓	✓	✓	✓	✓	✓	N.K
12/JUN/2026	11:35	✓	✓	✓	✓	✓	✓	✓	M.K
15/JUN/2026	08:19	✓	✓	✓	✓	✓	✓	✓	N.K
16/JUN/2026	14:10	✓	✓	✓	✓	✓	✓	✓	M.K
17/JUN/2026	9:37	✓	✓	✓	✓	✓	✓	✓	M.K
18/JUN/2026	8:18	✓	✓	✓	✓	✓	✓	✓	M.K
19/JUN/2026	9:53	✓	✓	✓	✓	✓	✓	✓	N.K
22/JUN/2026	12:15	✓	✓	✓	✓	✓	✓	✓	M.K
23/JUN/2026	12:16	✓	✓	✓	✓	✓	✓	✓	M.K
24/JUN/2026	09:56	✓	✓	✓	✓	✓	✓	✓	N.K
25/JUN/2026	13:16	✓	✓	✓	✓	✓	✓	✓	N.K
26/JUN/2026	09:02	✓	✓	✓	✓	異常音中	✓	✓	N.K
29/JUN/2026	09:17	✓	✓	✓	✓	✓	✓	✓	N.K
30/JUN/2026	9:57	✓	✓	✓	✓	✓	✓	✓	N.K

Ref-1 は 26/MAY/2026 11:15 ~ 03/JUN/2026 まで修理のため未使用。異常音は確認し、
09:00

03/JUN/2026 09:07より Ref-2 から Ref-1 に3台目集薬を戻し、運用再開した。 03/JUN/2026 Naoya Kazuyama

Ref-2 は 25/JUN/2026 13:30 ~ 29/JUN/2026 09:30 の間に冷蔵庫を交換した。